

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2018-05-29
KIT DESCRIPTION: IA PREMIUM PLUS LEVEL 1
CATALOGUE NO: IA3109
LOT NUMBERS: 1810EC
EXPIRY DATE: 2021-05-28
SYSTEM OF ANALYSIS: ROCHE COBAS E801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
ALPHAFOETOPROTEIN (KU/L)	1810EC	7.86-11.80	9.82	10.0	N/A
				9.5	
CEA (ug/l)	1810EC	3.30-4.96	4.13	3.90	N/A
				4.00	
TOTAL PSA (ug/l)	1810EC	1.51-2.51	2.01	1.84	N/A
				1.80	

QC TESTED BY: DAVINA GRAHAM
DATE: 12-NOV-18
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *C. Meutic*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NII 5738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016