



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2020-10-29
KIT DESCRIPTION:	Lipid Control 2
CATALOGUE NO:	LE2669
BATCH NO:	2835CH
EXPIRY DATE:	2023-10-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Level II/ 2835CH	HDL	1.35	1.15-1.55	1.34	mmol/L
Level II/ 2835CH	LDL	3.87	3.29-4.45	3.85	mmol/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) | [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MAn, R J Fitzgerald, W Fitzgerald OBE, R F Kely BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0904 00.