



HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2021-10
Product Description:	PRECISION CHEMISTRY CONTROL PREMIUM PLUS LEVEL 3
Catalogue Number:	UE1558
Batch Number:	1221UE
Expiry Date:	2025-06

Analyte	Lot No.	Target	Range		Result	Units
Total Protein	1221UE	44.64	N/A	N/A	47.017	g/L
Albumin	1221UE	30.93	N/A	N/A	30.630	g/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By:

Date:

21-OCT-21

Quality Control Representative

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
+44 (0) 28 9442 2413 | marketing@randox.com | randox.com

Directors: Dr S P FitzGerald CBE FREng DL BSc PhD DSc (Hons) FIAE Man. R P Kelly BSc FCA.

Randox Laboratories Ltd is a subsidiary of Randox Holdings Limited a company registered within Northern Ireland with company number NI. 614690. VAT Registered Number: GB 151 12748-QC

UE1558
Revision 1

Revision (0)
Last Printed Date:21/10/2021