

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-10-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 2124EC  
EXPIRY DATE: 2024-10-28  
SYSTEM OF ANALYSIS: ROCHE COBAS e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH (U/l)	2124EC	26.3 - 39.5	32.9	37.3	N/A
				37.1	
FT3 (pmol/l)	2124EC	11.3 - 18.9	15.1	15.3	N/A
				15.2	
FT4 (pmol/l)	2124EC	22.4 - 37.4	29.9	29.1	N/A
				29.7	

QC TESTED BY: DAVINA GRAHAM  
DATE: 09-Dec-21  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15733, VAT Number GB 353 0304 00.