

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-04-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1977EC
EXPIRY DATE: 2023-04-28
SYSTEM OF ANALYSIS: ROCHE COBAS 8000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Alpha-fetoprotein (KU/l)	1977EC	38.2 - 57.4	47.8	45.1	N/A
				43.8	
PSA Free (ng/ml)	1977EC	6.98 - 11.6	9.30	9.1	N/A
				9.1	
Insulin (mU/l)	1977EC	11.2 - 18.6	14.9	14.1	N/A
				14.1	

QC TESTED BY: AIMEE ANDERSON
DATE: 26-Feb-21
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016