

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE:			2020-1	10-29			
KIT DESCRIPTION:	_	IMMUNOA	SSAY	SPECIA	LITY II -	-LEVEL	1
CATALOGUE NO:	Ī.	AS3117					
LOT NUMBERS:	2017	'EC					
EXPIRY DATE:	2022	2-10-28					
SYSTEM OF ANALYSIS:		SIEMEN	SIMMU	JLITE 20	00		

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Gastrin (pmol/l)	2017EC	13.3 - 22.1	17.7	15.90	N/A
				16.30	
Calcitonin (pg/ml)	2017EC	19.6 - 32.6	26.1	28.30	N/A
	***************************************			28.00	
		- ALAVIII WILLIAM			
-					

QC TESTED BY:		Charlene Junkin		
DATE:	24-Mar-21			
RESULT	(PASS/FAIL):	PASS .		
QC APPI	ROVED BY:	a. Mashi		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr. S. P. Fitzgerald, BSc. PhD. DSc. (Hons). FIAE Man., R. J. Fitzgerald, W. Fitzgerald, OBE, R. P. Kelly BSc. PCA, Randox Laboratories Limited is a company registered in Northern Ireland, Company Number, NH 5733, VAT. Number, GB, 353, 0304, 00.