

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2020-11-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY I - LEVEL 1
CATALOGUE NO: IAS3113
LOT NUMBERS: 2012EC
EXPIRY DATE: 2022-11-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Anti TG (kU/l)	2012EC	22.1 - 36.9	29.5	33.7	N/A
				33.6	
Anti TPO (kU/l)	2012EC	35.6 - 59.4	47.5	43.7	N/A
				41.2	

QC TESTED BY: Charlene Junkin
DATE: 11-MAR-21
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *G. Mentis*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA,
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016