

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-03-29
KIT DESCRIPTION: Immunoassay Speciality 1 – Level 1
CATALOGUE NO: IAS3113
LOT NUMBERS: 2064EC
EXPIRY DATE: 2023-03-28
SYSTEM OF ANALYSIS: Roche Cobas 8000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
25-OH Vitamin D (nmol/l)	2064EC	22.0 - 36.6	29.3	34	N/A
				30	
Insulin (mU/l)	2064EC	3.44 - 5.74	4.59	3.96	N/A
				3.87	

QC TESTED BY: Davina Graham
DATE: 08-DEC-21
RESULT (PASS/FAIL): PASS
QC APPROVED BY: G. Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.