

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-12-29
KIT DESCRIPTION: Immunoassay Premium Plus – Level 2
CATALOGUE NO: IA3110
LOT NUMBERS: 2107EC
EXPIRY DATE: 2024-12-28
SYSTEM OF ANALYSIS: Roche Cobas_8000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FT3 (pmol/l)	2107EC	10.2 - 17.0	13.6	13.5	N/A
				13.3	
FT4 (pmol/l)	2107EC	21.9 - 36.5	29.2	30.1	N/A
				29.2	
TSH (mIU/l)	2107EC	1.76 - 2.64	2.20	2.12	N/A
				2.13	

QC TESTED BY: NAOMI CLYDE
DATE: 14-APR-22
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *A. Murtan*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com
Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FHE Man, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number: NI15738, VAT Number: GB 353 0304 00.