

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-10-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 3  
CATALOGUE NO: IA3111  
LOT NUMBERS: 2079EC  
EXPIRY DATE: 2024-10-28  
SYSTEM OF ANALYSIS: ROCHE COBAS e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH(U/L)	2079EC	44.4 - 66.6	55.5	62.3	N/A
				63.3	
FT3 (pmol/l)	2079EC	22.4 - 37.3	29.8	29.7	N/A
				29.9	
FT4 (pmol/l)	2079EC	54.2 - 90.4	72.3	72.7	N/A
				71.3	

QC TESTED BY: Davina Graham  
DATE: 09-DEC-21  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016