

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-03-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II – LEVEL 3
CATALOGUE NO: IAS3119
LOT NUMBERS: 2174EC
EXPIRY DATE: 2024-03-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	2174EC	239 – 397	318	359	N/A
				355	
Gastrin (pmol/l)	2174EC	143 – 239	191	207	N/A
				210	

QC TESTED BY: Kelly Jennings
DATE: 14-Sep-22
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *R. Mearns*

Randox Laboratories Limited, 53 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0)-28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Med. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA, Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI 5736. VAT Number GB 353 0304 00.