

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2021-10-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS TRI LEVEL  
CATALOGUE NO: IA3112  
LOT NUMBERS: 603640 ( 2105EC, 2107EC, 2147EC)  
EXPIRY DATE: 2024-10-28  
SYSTEM OF ANALYSIS: Roche Cobas\_8000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH (mU/ml)	2105EC	5.52 - 8.28	6.90	7.10	N/A
				7.00	
	2107EC	20.3 – 30.5	25.4	26.0	N/A
				25.8	
	2147EC	44.4 – 66.6	55.5	56.4	N/A
				56.1	
LH (mU/ml)	2105EC	2.86 – 4.30	3.58	3.70	N/A
				3.70	
	2107EC	27.7 – 41.5	34.6	34.7	N/A
				34.9	
	2147EC	43.8 – 65.8	54.8	55.8	N/A
				55.0	
Progesterone (nmol/l)	2105EC	2.61 – 3.91	3.26	3.45	N/A
				3.45	
	2107EC	26.4 – 39.6	33.0	32.6	N/A
				33.2	
	2147EC	81.6 - 122	102	91.7	N/A
				97.5	

### QC TESTED BY:

Davina Graham

DATE: 08-JUN-22

RESULT (PASS/FAIL): PASS

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016

# RANDOX

QC APPROVED BY:

A. Martin

**Randox Laboratories Limited**, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
**T** +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** [marketing@randox.com](mailto:marketing@randox.com) **I** [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number: NI: 5738. VAT Number: GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016