RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-05-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY 1 – LEVEL 3
CATALOGUE NO: IAS3115
LOT NUMBERS: 2195EC
EXPIRY DATE: 2024-05-28
SYSTEM OF ANALYSIS: ROCHE COBAS 8000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
IGF 1 (ug/l)	2195EC	93 - 155	124	129.3	N/A
				130.0	
Parathyroid Hormone (pmol/l)	2195EC	59.0 – 98.3	78.6	81.7	N/A
				82.6	

QC TESTED BY:
Davina Graham
DATE: 29-Jul-22
RESULT (PASS/FAIL): PASS
QC APPROVED BY: A. Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. P P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

RANDOX

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number: NI: 5738. VAT Number: GB 353 0304 00.