



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

|                      |                 |
|----------------------|-----------------|
| DATE OF MANUFACTURE: | 2022-03-29      |
| KIT DESCRIPTION:     | LIPID CONTROL 2 |
| CATALOGUE NO:        | LE2669          |
| BATCH NO:            | 2969CH          |
| EXPIRY DATE:         | 2025-03-28      |

| SAMPLE            | ANALYTE | TARGET | RANGE     | RESULT | UNITS  |
|-------------------|---------|--------|-----------|--------|--------|
| Level II / 2969CH | LDL     | 3.72   | 3.16-4.28 | 3.57   | mmol/L |
| Level II / 2969CH | HDL     | 1.44   | 1.22-1.66 | 1.42   | mmol/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRCM, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15739, VAT Number GB 353 0304 00.