

RANDOX

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2022-10
KIT DESCRIPTION:	ALDOLASE CONTROL LEVEL 3
CATALOGUE NO:	AD55002
BATCH NO:	479AD
EXPIRY DATE:	2025-09

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
Aldolase Control Level 3	Aldolase	20.2	15.2-25.2	19.1	U/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FRC Path,
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15739, VAT Number GB 353 0304 00.