

RANDOX

CERTIFICATE OF ANALYSIS


This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-06-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2
CATALOGUE NO: IA3110
LOT NUMBERS: 2228EC
EXPIRY DATE: 2025-06-28
SYSTEM OF ANALYSIS: Roche Cobas_e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH (mU/ml)	2228EC	22.1 - 33.1	27.6	28.6	N/A
				28.3	
LH (mU/ml)	2228EC	27.8 - 41.8	34.8	34.3	N/A
				34.3	
Progesterone (nmol/l)	2228EC	26.3 - 39.5	32.9	31.6	N/A
				31.5	

QC TESTED BY: Davina Graham
DATE: 11-Oct-22
RESULT (PASS/FAIL): PASS
QC APPROVED BY: A. Martin



Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S.P. Fitzgerald BSc PhD DSc (Hons) FIAE Man R.J. Fitzgerald, W. Fitzgerald OBE, R.P. Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016