



HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2021-12-29
Product Description:	HUMAN ASSAYED MULTI-SERA
Catalogue Number:	HN1530
Batch Number:	1585UN
Expiry Date:	2025-12-28

Analyte	Lot No.	Target	Range		Result	Units
ALBUMIN	1585UN	41.60	35.40	47.80	41.02	g/L
TOTAL PROTEIN	1585UN	60.70	48.50	72.90	61.74	g/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By:

Date:

07-JUN-23

Quality Control Representative