

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-06-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 2180EC  
EXPIRY DATE: 2025-06-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Insulin ( $\mu$ U/ml)	2180EC	12.1 - 20.1	16.1	15.7	N/A
				15.4	
SHBG (nmol/l)	2180EC	26.1 - 39.1	32.6	32.4	N/A
				31.9	
Testosterone (nmol/l)	2180EC	11.5 - 17.3	14.4	14.6	N/A
				14.4	

### QC TESTED BY:

Kelly Jennings

DATE: 06-Mar-23

RESULT (PASS/FAIL): \_\_\_\_\_

QC APPROVED BY: *G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number: NII 5738, VAT Number: GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016