

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a  
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-11-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS – LEVEL 3  
CATALOGUE NO: IA3111  
LOT NUMBERS: 2231EC  
EXPIRY DATE: 2025-11-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
FSH (mU/ml)	2231EC	45.9 – 68.9	57.4	55.7	N/A
				54.8	
LH (mU/ml)	2231EC	42.9 – 64.3	53.6	53.7	N/A
				53.6	
Cortisol (nmol/l)	2231EC	769 – 1281	1025	1174	N/A
				1184	

QC TESTED BY: Davina Graham  
DATE: 16-May-23  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *G. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man R J Fitzgerald, W Fitzgerald OBE, R P Kelly BS: FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number: NI15758, VAT Number: GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016