

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-10-29
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II – LEVEL 3
CATALOGUE NO: IAS3119
LOT NUMBERS: 2225EC
EXPIRY DATE: 2024-10-28
SYSTEM OF ANALYSIS: SIEMENS IMMULITE 2000 & ROCHE COBAS e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	2225EC	358 - 596	477	588	N/A
				572	
Gastrin (pmol/l)	2225EC	155 – 258	206	225	N/A
				224	

QC TESTED BY: Davina Graham
DATE: 31-Mar-23
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *E. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number: NI15738. VAT Number: GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016