



# RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2022-04-29</b>
<b>KIT DESCRIPTION:</b>	<b>HAEMOGLOBIN A<sub>1c</sub> CONTROL SET (HbA<sub>1c</sub> CONTROL)</b>
<b>CATALOGUE NO:</b>	<b>HA5072</b>
<b>BATCH NO:</b>	<b>640602</b>
<b>EXPIRY DATE:</b>	<b>2025-04-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>LEVEL I / 2341HA</b>	<b>HbA<sub>1c</sub></b>	<b>5.14</b>	<b>4.11 - 6.17</b>	<b>5.60</b>	<b>% HbA<sub>1c</sub></b>
<b>LEVEL II / 2343HA</b>	<b>HbA<sub>1c</sub></b>	<b>13.90</b>	<b>11.08 - 16.60</b>	<b>13.29</b>	<b>% HbA<sub>1c</sub></b>

APPROVED BY:

01-AUG-23

## QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S. P. Fitzgerald BSc PhD DSc (Hons) FIAE Man. R.J. Fitzgerald W. Fitzgerald OBE R.P. Kelly BSc FCA  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.