

RANDOX

HUMAN SERUM QUALITY CONTROL

CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2023-03-29
Product Description:	ASSAYED HUMAN MULTI-SERA
Catalogue Number:	HE1532
Batch Number:	1328UE
Expiry Date:	2027-03-28

Analyte	Lot No.	Target	Range		Result	Units
ALBUMIN	1328UE	30.40	25.80	35.00	29.83	g/L
TOTAL PROTEIN	1328UE	47.50	38.00	57.00	47.48	g/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By:

Maria Kane

SignNow e-signature ID: bc1a283ccf...
12-Sep-2023 09:39:23 UTC
Maria Kane (Signer)

Date:

12-SEP-2023

Quality Control Representative