

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-10-29  
KIT DESCRIPTION: IMMUNOASSAY SPECIALITY II LEV 1  
CATALOGUE NO: IAS3117  
LOT NUMBERS: 2223EC  
EXPIRY DATE: 2024-10-28  
SYSTEM OF ANALYSIS: Roche Cobas e801 and Siemens Immulite 2000

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Calcitonin (pg/ml)	2223EC	16.7 – 27.9	22.3	24.9	N/A
				24.7	
Gastrin (pmol/l)	2223EC	7.30 – 12.2	9.73	10.8	N/A
				10.1	

### QC TESTED BY:

Hannah Anglin

DATE: 28-Feb-23

RESULT (PASS/FAIL): PASS

QC APPROVED BY: G. Mantu

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016