



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2023-03-29
Product Description:	Liquid Protein Calibrator Series
Catalogue Number:	IT2691
Batch Number:	647343
Expiry Date:	2025-03-28

Analyte	Lot No.	Target	Range		Result	Units
C4	648LPC	36.2	29	43.4	35.000	mg/dL
C4	602LPC	42.1	33.7	50.5	40.950	mg/dL

Approved By:

SignNow e-signature ID: d33be6062...
15/09/2023 14:51:48 UTC
Ciara Mc Guigan (Signer)

Date:

15/09/2023

Quality Control Representative

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
+44 (0) 28 9442 2413 | marketing@randox.com | randox.com

Directors: Dr S P FitzGerald CBE FREng DL BSc PhD DSc (Hons) FIAE Man. R P Kelly BSc FCA.
Randox Laboratories Ltd is a subsidiary of Randox Holdings Limited a company registered within Northern Ireland with company number NI. 614690. VAT Registered Number: GB 151 6827 08.