

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2023-06-29
KIT DESCRIPTION:	TRI LEVEL CARDIAC CONTROL (CRD CONTROL 1,2,3)
CATALOGUE NO:	CQ3259
BATCH / LOT NO:	659706
EXPIRY DATE:	2027-06-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 4783CK	CK NAC	87.0	71.0-103.0	89.6	U/L
LEVEL II / 4784CK	CK NAC	209	171-247	230	U/L
LEVEL III / 4785CK	CK NAC	597	490-704	633	U/L
LEVEL II / 4784CK	CK MB	22.5	18.0-27.0	22.6	U/L
LEVEL III / 4685CK	CK MB	93.4	74.7-112.0	98.0	U/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald CBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.