

# RANDOX

## CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-11-29  
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 2  
CATALOGUE NO: IA3110  
LOT NUMBERS: 2230EC  
EXPIRY DATE: 2025-11-28  
SYSTEM OF ANALYSIS: Roche Cobas e801

### TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA125 (U/ml)	2230EC	62.6 – 94.0	78.3	76	N/A
				77	
CA15-3 (U/ml)	2230EC	33.8 – 50.8	42.3	42	N/A
				43	
CA19-9 (U/ml)	2230EC	62.2 – 93.2	77.7	76.3	N/A
				75.9	

QC TESTED BY: Davina Graham  
DATE: 05-Jul-23  
RESULT (PASS/FAIL): PASS  
QC APPROVED BY: *R. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA,  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15733. VAT Number GB 353 0304 00.

FORM NO. 6521-QC  
REVISION (4)  
14 April 2016