

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: 2022-11-29
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS - LEVEL 3
CATALOGUE NO: IA3111
LOT NUMBERS: 2292EC
EXPIRY DATE: 2025-11-28
SYSTEM OF ANALYSIS: Roche Cobas e801

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CA125 (U/ml)	2292EC	138 – 208	173	170	N/A
				172	
AFP (KIU/l)	2292EC	158 – 238	198	180	N/A
				181	
CA19-9 (U/ml)	2292EC	130 – 194	162	165	N/A
				166	

QC TESTED BY:

Andrew Gilmore

DATE: 07-Jul-23

RESULT (PASS/FAIL): PASS

QC APPROVED BY: *A. Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (H-Long) FIAE Man. R J Fitzgerald. W Fitzgerald OBE. R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15738. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016