

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

<b>DATE OF MANUFACTURE:</b>	<b>2023-06-29</b>
<b>KIT DESCRIPTION:</b>	<b>LIQUID ASSAYED SPECIFIC PROTEIN CONTROL (SP CONTROL)</b>
<b>CATALOGUE NO:</b>	<b>PS2682</b>
<b>BATCH NO:</b>	<b>675LPC</b>
<b>EXPIRY DATE:</b>	<b>2025-06-28</b>

<b>SAMPLE</b>	<b>ANALYTE</b>	<b>TARGET</b>	<b>RANGE</b>	<b>RESULT</b>	<b>UNITS</b>
<b>Level I / 675LPC</b>	<b>Albumin</b>	<b>20.3</b>	<b>17.3 – 23.3</b>	<b>20.7</b>	<b>g/L</b>
<b>Level I / 675LPC</b>	<b>Total Protein</b>	<b>38.8</b>	<b>31.0-46.6</b>	<b>39.7</b>	<b>g/L</b>

APPROVED BY:



QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom  
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E [marketing@randox.com](mailto:marketing@randox.com) I [www.randox.com](http://www.randox.com)

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAC Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.  
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.