



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2021-09-29
KIT DESCRIPTION:	BLOOD GAS CONTROL LEVEL 3
CATALOGUE NO:	BG5003
BATCH NO:	252BG
EXPIRY DATE:	2023-09-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL III / 252BG	POTASSIUM	6.15	5.66 – 6.64	6.27	mmol/L
LEVEL III / 252BG	SODIUM	165.00	157.00 – 173.00	167.00	mmol/L

APPROVED BY:

H Martin

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRAE Man. R J Fitzgerald W Fitzgerald OBE R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15758, VAT Number GB 353 0304 00