

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-04
KIT DESCRIPTION:	C-REACTIVE PROTEIN FULL RANGE CALIBRATOR SERIES
CATALOGUE NO:	CP2499
BATCH NO:	383468
EXPIRY DATE:	2017-08

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 402LPC	48.4	38.7-58.1	47.0	mg/L
LEVEL III / 403LPC	72.5	58.0-87.0	71.8	mg/L

APPROVED BY:

vette Milla-

QUALITY CONTROL REPRESENTATIVE