

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-09
KIT DESCRIPTION:	LIQUID PROTEIN CALIBRATOR SERIES
CATALOGUE NO:	IT2691
BATCH NO:	396389
EXPIRY DATE:	2018-01-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 454LPC	HAPTOGLOBIN	1.00	0.80-1.20	0.98	g/L
LEVEL III / 455LPC	HAPTOGLOBIN	1.50	1.20-1.80	1.52	g/L

APPROVED BY:

Juste Milla-

QUALITY CONTROL REPRESENTATIVE