

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-09
KIT DESCRIPTION:	HbA1c CALIBRATOR
CATALOGUE NO:	HA3444
BATCH NO:	396816
EXPIRY DATE:	2018-03

SAMPLE	TARGET	RANGE	RESULT	UNITS
HbA1c CONTROL LEVEL I	5.8	4.64-6.96	6.07	%HbA1c
HbA1c CONTROL LEVEL II	12.4	9.92-14.9	12.62	%HbA1c

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com