



RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04
KIT DESCRIPTION:	CRP Calibrator
CATALOGUE NO:	CP2179
BATCH NO:	2476CP
EXPIRY DATE:	2018-09-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
CRP Control / Level II 2453CP	22.0	17.6-26.4	21.3	mg/L
CRP Control / Level III 2454CP	147.6	118.1-177.1	147.4	mg/L

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R F Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI5738, VAT Number GB 353 0304 00.

4706-QC Rev4
27th September 2013