



HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-02
KIT DESCRIPTION:	HUMAN ELEVATED SERUM CONTROL III
CATALOGUE NO:	HE1532
BATCH / LOT NO:	884UE
EXPIRY DATE:	2020-08

ANALYTE	TARGET	RANGE	RESULT	UNITS
ALB	30.1	25.5 – 34.7	29.6	g/L
TP	46.5	37.2 – 55.8	46.3	g/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FRC Path, R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI5738, VAT Number GB 353 0304 00.