

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-02-28
KIT DESCRIPTION:	HIGH SENSITIVITY CRP CALIBRATOR SERIES
CATALOGUE NO:	CP2478
BATCH NO:	412228
EXPIRY DATE:	2018-02-28

SAMPLE	TARGET	RANGE	RESULT	UNITS
LEVEL I / 2429CP	1.03	0.77-1.29	1.02	mg/L
LEVEL II / 2430CP	4.34	3.47-5.21	4.33	mg/L

APPROVED BY:

Yvette Millar

QUALITY CONTROL REPRESENTATIVE