

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2015-04
KIT DESCRIPTION:	CK-MB CONTROL
CATALOGUE NO:	CK1212
BATCH NO:	3821CK
EXPIRY DATE:	2019-03

ANALYTE	TARGET	RANGE	RESULT	UNITS
СК-МВ	158	134-182	147	mmol/L
CK-NAC	248	211-285	250	mmol/L

**APPROVED BY:** 

Eileen Beagan

**QUALITY CONTROL REPRESENTATIVE**