

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-09
KIT DESCRIPTION:	H-FABP CONTROL LEVEL 2
CATALOGUE NO:	FB4027
BATCH NO:	128FB
EXPIRY DATE:	2018-08

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II/128FB	H-FABP	29.33	23.46-35.20	31.50	ng/ml

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE