

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-12
KIT DESCRIPTION:	BLOOD GAS CONTROL 3
CATALOGUE NO:	BG5003
BATCH NO:	165BG
EXPIRY DATE:	2017-11-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
CONTROL 1/ 163BG	POTASSIUM	6.32	5.81-6.83	6.27	mmol/l
CONTROL 1/ 163BG	SODIUM	161.00	153.00-169.00	162.40	mmol/l

APPROVED BY:

Grainne Tool

QUALITY CONTROL REPRESENTATIVE