

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-01
KIT DESCRIPTION:	COAGULATION CONTROL 3
CATALOGUE NO:	CG5023
BATCH NO:	182CG
EXPIRY DATE:	2018-03

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
182CG	PROTHROMBIN TIME	50.5	40.4-60.6	52.9	secs
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

vette Milla-