

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-04	
KIT DESCRIPTION:	FRUCTOSAMINE CONTROL HIGH	
CATALOGUE NO:	FR2996	
BATCH NO:	477FR	
EXPIRY DATE:	2018-03	

SAMPLE	TARGET	RANGE	RESULT	UNITS
477FR	794	635 - 953	768	µmol/L
N/A	N/A	N/A	N/A	N/A

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | www.randox.com