

HUMAN SERUM QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-07
KIT DESCRIPTION:	CARDIAC CONTROL SET
CATALOGUE NO:	CQ3259
BATCH / LOT NO:	389825
EXPIRY DATE:	2019-06

CONTROL	ANALYTE	TARGET	RANGE	RESULT	UNITS
3845CK	CK-NAC	74	61-87	74	U/L
3846CK	CK-NAC	247	203-291	244	U/L
3847CK	CK-NAC	552	453-651	548	U/L
3846CK	СК-МВ	25.0	20.0-30.0	26.7	U/L
3847CK	СК-МВ	134	107-161	139	U/L

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

APPROVED BY:

S Mc Guern

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgenald BSc PhD DSc (Hons) FIAE Man. R J Fitzgenald. W Fitzgenald OBE. R P Ketly BSc FCA. Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NII5738. VAT Number GB 353 0304 00.