

糖化血红蛋白 (HbA_{1c}) 质控品

货号: HA5072 包装: 2x2x0.5 ml
批号: 430317 (1898HA & 效期: 2020-04
1901HA)

产品描述

本质控品适用于临床化学体外诊断系统检测糖化血红蛋白 (HbA_{1c}) 的质量控制。朗道供应两个水平: 低浓度水平 (水平 1) 和正常浓度水平 (水平 2)。

安全预防措施和警告

本产品仅用于体外诊断。禁止用口吸。按照实验室常规预防措施对试剂进行处理。

警告: 本品为潜在的生物危险物质。

该质控品采用人血, 对所有捐献者的血液均进行了人类免疫缺陷病毒 HIV (HIV1、HIV2) 抗体、乙型肝炎表面抗原 (HbsAg) 和丙型肝炎病毒 (HCV) 抗体的测试, 发现均呈阴性。所采用的方法均经 FDA 认证。

但既然没有一种方法能够完全保证其没有传染物质, 因此该质控品和所有的病人样品均应当按照能够传播疾病的样品小心处理。

保存和稳定性

未开瓶, 2~8°C 可保存至效期末。

复溶后, 2~8°C 保存可稳定 1 个月。稳定性数据基于免疫比浊法方法学实验数据的统计。

注意: 复溶后不要冷冻。

使用说明

该糖化血红蛋白质控品为冻干品。

1. 小心打开瓶盖;
2. 准确量取 0.5 ml 的蒸馏水加入;
3. 盖好瓶盖, 转动试剂瓶若干次, 室温下静置 15 分钟;
4. 15 分钟后, 旋转和翻转试剂瓶, 继续复溶, 直到所有的冻干材料已溶解及溶液已混合均匀为止。

注意: 按照所使用试剂盒的说明书, 应将质控品按照与病人样本相同的方式处理。如果使用朗道试剂盒分析糖化血红蛋白和总血红蛋白, 须进行预处理: 将 10 µL 质控液与 400 µL 血红蛋白变性试剂混合作 1:41 稀释。

提供的材料

浓度水平	水平 1 (HbA _{1c} 质控 1)	水平 2 (HbA _{1c} 质控 2)
包装规格	2 x 0.5 ml	2 x 0.5 ml

需要自备的材料

移液管, 蒸馏水

赋值

每一批质控血液都要送到大量的外部实验室, 通过对这些实验室返回的结果统计而赋值。

注: 详细赋值信息请以原版英文说明书为准, 原版说明书请在英国朗道公司官网 www.randox.com 进行下载。

LEVEL I

Method	Units	Target	Range
Abbott Architect c systems (DCCT / NGSP)	%HbA _{1c}	5.25	4.20 – 6.30
Abbott Architect c systems (IFCC)	mmol/mol	33.9	27.1 – 40.7
Abbott Architect c (Direct Turbidimetric) (DCCT / NGSP)	%HbA _{1c}	6.05	4.84 – 7.26
Abbott Architect c (Direct Turbidimetric) (IFCC)	mmol/mol	42.6	34.1 – 51.1
Abbott Architect i systems (DCCT / NGSP)	%HbA _{1c}	6.43	5.14 – 7.72
Abbott Architect i systems (IFCC)	mmol/mol	46.8	37.4 – 56.2
Arkray Menarini HA8121/40/60/80 (DCCT / NGSP)	%HbA _{1c}	5.68	4.54 – 6.82
Arkray Menarini HA8121/40/60/80 (IFCC)	mmol/mol	38.6	30.9 – 46.3
Beckman AU400/480/600/640/2700/5400 (DCCT / NGSP)	%HbA _{1c}	5.85	4.68 – 7.02
Beckman AU400/480/600/640/2700/5400 (IFCC)	mmol/mol	40.4	32.3 – 48.5
Beckman DxC600/DxC800 (DCCT / NGSP)	%HbA _{1c}	5.75	4.60 – 6.90
Beckman DxC600/DxC800 (IFCC)	mmol/mol	39.3	31.4 – 47.2
Bioanalytic Diagnostic HbA _{1c} (DCCT / NGSP)	%HbA _{1c}	6.00	4.80 – 7.20
Bioanalytic Diagnostic HbA _{1c} (IFCC)	mmol/mol	42.1	33.7 – 50.5
Biorad D-10 (DCCT / NGSP)	%HbA _{1c}	5.68	4.54 – 6.82
Biorad D-10 (IFCC)	mmol/mol	38.6	30.9 – 46.3
Biorad Variant II (ion exchange) (DCCT / NGSP)	%HbA _{1c}	5.72	4.58 – 6.86
Biorad Variant II (ion exchange) (IFCC)	mmol/mol	39.0	31.2 – 46.8
EKF Quo-Lab A1c (DCCT / NGSP)	%HbA _{1c}	6.60	5.28 – 7.92
EKF Quo-Lab A1c (IFCC)	mmol/mol	48.6	38.9 – 58.3
Erba XL Series (DCCT / NGSP)	%HbA _{1c}	5.77	4.62 – 6.92
Erba XL Series (IFCC)	mmol/mol	39.6	31.7 – 47.5
Konelab 20/30/60 / Thermo Indiko (DCCT / NGSP)	%HbA _{1c}	6.16	4.93 – 7.39
Konelab 20/30/60 / Thermo Indiko (IFCC)	mmol/mol	43.8	35.0 – 52.6
Mindray BS200/300/400 (DCCT / NGSP)	%HbA _{1c}	5.60	4.48 – 6.72
Mindray BS200/300/400 (IFCC)	mmol/mol	37.7	30.2 – 45.2
Ortho Vitros 4600 / 5600 / 5.1 FS (DCCT / NGSP)	%HbA _{1c}	5.78	4.62 – 6.94
Ortho Vitros 4600 / 5600 / 5.1 FS (IFCC)	mmol/mol	39.7	31.8 – 47.6
Randox Rx Series (DCCT / NGSP)	%HbA _{1c}	5.96	4.77 – 7.15
Randox Rx Series (IFCC)	mmol/mol	41.6	33.3 – 49.9
Roche Cobas 4000/c311 (DCCT / NGSP)	%HbA _{1c}	5.63	4.50 – 6.76
Roche Cobas 4000/c311 (IFCC)	mmol/mol	38.0	30.4 – 45.6
Roche Cobas 6000/8000 (DCCT / NGSP)	%HbA _{1c}	5.62	4.50 – 6.74
Roche Cobas 6000/8000 (IFCC)	mmol/mol	37.9	30.3 – 45.5
Roche Cobas c513 (DCCT / NGSP)	%HbA _{1c}	5.67	4.54 – 6.80
Roche Cobas c513 (IFCC)	mmol/mol	38.5	30.8 – 46.2
Roche Integra (DCCT / NGSP)	%HbA _{1c}	5.70	4.56 – 6.84
Roche Integra (IFCC)	mmol/mol	38.8	31.0 – 46.6
Roche Modular P/Cobas c111 (DCCT / NGSP)	%HbA _{1c}	6.11	4.89 – 7.33
Roche Modular P/Cobas c111 (IFCC)	mmol/mol	43.3	34.6 – 52.0
Sebia Capillarys / Minicap (DCCT / NGSP)	%HbA _{1c}	5.52	4.42 – 6.62
Sebia Capillarys / Minicap (IFCC)	mmol/mol	36.8	29.4 – 44.2
Siemens ADVIA 1200/1650/1800/2400 (DCCT / NGSP)	%HbA _{1c}	5.83	4.66 – 7.00
Siemens ADVIA 1200/1650/1800/2400 (IFCC)	mmol/mol	40.2	32.2 – 48.2
Siemens DCA2000/Vantage (DCCT / NGSP)	%HbA _{1c}	6.21	4.97 – 7.45
Siemens DCA2000/Vantage (IFCC)	mmol/mol	44.4	35.5 – 53.3

LEVEL 1 (continued)

Method	Units	Target	Range
Siemens/Dade Dimension (DCCT / NGSP)	%HbA _{1c}	6.06	4.85 – 7.27
Siemens/Dade Dimension (IFCC)	mmol/mol	42.7	34.2 – 51.2
Tokyo Boeki/Prestige 24i (DCCT / NGSP)	%HbA _{1c}	5.75	4.60 – 6.90
Tokyo Boeki/Prestige 24i (IFCC)	mmol/mol	39.3	31.4 – 47.2
TOSOH HLC723/G7/G8/GX (DCCT / NGSP)	%HbA _{1c}	5.66	4.53 – 6.79
TOSOH HLC723/G7/G8/GX (IFCC)	mmol/mol	38.4	30.7 – 46.1
Trin Bio CLC385/PDQ/Ultra 2 (DCCT / NGSP)	%HbA _{1c}	6.13	4.90 – 7.36
Trin Bio CLC385/PDQ/Ultra 2 (IFCC)	mmol/mol	43.5	34.8 – 52.2
Trinity Biotech Tri-stat (DCCT / NGSP)	%HbA _{1c}	6.91	5.53 – 8.29
Trinity Biotech Tri-stat (IFCC)	mmol/mol	52.0	41.6 – 62.4
Trinity/Menarini Premier Hb9210 (DCCT / NGSP)	%HbA _{1c}	6.05	4.84 – 7.26
Trinity/Menarini Premier Hb9210 (IFCC)	mmol/mol	42.6	34.1 – 51.1

Total Haemoglobin		Target	Range
Randox Rx Series	g/dl	12.8	10.2 – 15.4

LEVEL 2

Method	Units	Target	Range
Abbott Architect c systems (DCCT / NGSP)	%HbA _{1c}	10.6	8.48 – 12.7
Abbott Architect c systems (IFCC)	mmol/mol	92.4	73.9 – 111
Abbott Architect c (Direct Turbidimetric) (DCCT / NGSP)	%HbA _{1c}	11.1	8.88 – 13.3
Abbott Architect c (Direct Turbidimetric) (IFCC)	mmol/mol	97.8	78.2 – 117
Abbott Architect i systems (DCCT / NGSP)	%HbA _{1c}	13.2	10.6 – 15.8
Abbott Architect i systems (IFCC)	mmol/mol	121	96.8 – 145
Arkray Menarini HA8121/40/60/80 (DCCT / NGSP)	%HbA _{1c}	10.9	8.72 – 13.1
Arkray Menarini HA8121/40/60/80 (IFCC)	mmol/mol	95.6	76.5 – 115
Beckman AU400/480/600/640/2700/5400 (DCCT / NGSP)	%HbA _{1c}	11.4	9.12 – 13.7
Beckman AU400/480/600/640/2700/5400 (IFCC)	mmol/mol	101	80.8 – 121
Beckman DxC600/DxC800 (DCCT / NGSP)	%HbA _{1c}	11.7	9.36 – 14.0
Beckman DxC600/DxC800 (IFCC)	mmol/mol	104	83.2 – 125
Bioanalytic Diagnostic HbA _{1c} (DCCT / NGSP)	%HbA _{1c}	11.9	9.52 – 14.3
Bioanalytic Diagnostic HbA _{1c} (IFCC)	mmol/mol	107	85.6 – 128
Biorad D-10 (DCCT / NGSP)	%HbA _{1c}	10.9	8.72 – 13.1
Biorad D-10 (IFCC)	mmol/mol	95.6	76.5 – 115
Biorad Variant II (ion exchange) (DCCT / NGSP)	%HbA _{1c}	11.0	8.80 – 13.2
Biorad Variant II (ion exchange) (IFCC)	mmol/mol	96.7	77.4 – 116
EKF Quo-Lab A1c (DCCT / NGSP)	%HbA _{1c}	11.6	9.28 – 13.9
EKF Quo-Lab A1c (IFCC)	mmol/mol	103	82.4 – 124
Erba XL Series (DCCT / NGSP)	%HbA _{1c}	11.6	9.28 – 13.9
Erba XL Series (IFCC)	mmol/mol	103	82.4 – 124
Konelab 20/30/60 / Thermo Indiko (DCCT / NGSP)	%HbA _{1c}	11.4	9.12 – 13.7
Konelab 20/30/60 / Thermo Indiko (IFCC)	mmol/mol	101	80.8 – 121
Mindray BS200/300/400 (DCCT / NGSP)	%HbA _{1c}	11.3	9.04 – 13.6
Mindray BS200/300/400 (IFCC)	mmol/mol	100	80.0 – 120

LEVEL 2 (continued)

Method	Units	Target	Range
Ortho Vitros 4600 / 5600 / 5.1 FS (DCCT / NGSP)	%HbA _{1c}	11.4	9.12 – 13.7
Ortho Vitros 4600 / 5600 / 5.1 FS (IFCC)	mmol/mol	101	80.8 – 121
Randox Rx Series (DCCT / NGSP)	%HbA _{1c}	12.2	9.76 – 14.6
Randox Rx Series (IFCC)	mmol/mol	110	88.0 – 132
Roche Cobas 4000/c311 (DCCT / NGSP)	%HbA _{1c}	11.3	9.04 – 13.6
Roche Cobas 4000/c311 (IFCC)	mmol/mol	100	80.0 – 120
Roche Cobas 6000/8000 (DCCT / NGSP)	%HbA _{1c}	11.2	8.96 – 13.4
Roche Cobas 6000/8000 (IFCC)	mmol/mol	98.9	79.1 – 119
Roche Cobas c513 (DCCT / NGSP)	%HbA _{1c}	11.4	9.12 – 13.7
Roche Cobas c513 (IFCC)	mmol/mol	101	80.8 – 121
Roche Integra (DCCT / NGSP)	%HbA _{1c}	11.3	9.04 – 13.6
Roche Integra (IFCC)	mmol/mol	100	80.0 – 120
Roche Modular P/Cobas c111 (DCCT / NGSP)	%HbA _{1c}	11.6	9.28 – 13.9
Roche Modular P/Cobas c111 (IFCC)	mmol/mol	103	82.4 – 124
Sebia Capillarys / Minicap (DCCT / NGSP)	%HbA _{1c}	10.7	8.56 – 12.8
Sebia Capillarys / Minicap (IFCC)	mmol/mol	93.4	74.7 – 112
Siemens ADVIA 1200/1650/1800/2400 (DCCT / NGSP)	%HbA _{1c}	11.1	8.88 – 13.3
Siemens ADVIA 1200/1650/1800/2400 (IFCC)	mmol/mol	97.8	78.2 – 117
Siemens DCA2000/Vantage (DCCT / NGSP)	%HbA _{1c}	12.0	9.60 – 14.4
Siemens DCA2000/Vantage (IFCC)	mmol/mol	108	86.4 – 130
Siemens/Dade Dimension (DCCT / NGSP)	%HbA _{1c}	11.3	9.04 – 13.6
Siemens/Dade Dimension (IFCC)	mmol/mol	100	80.0 – 120
Tokyo Boeki/Prestige 24i (DCCT / NGSP)	%HbA _{1c}	11.0	8.80 – 13.2
Tokyo Boeki/Prestige 24i (IFCC)	mmol/mol	96.7	77.4 – 116
TOSOH HLC723/G7/G8/GX (DCCT / NGSP)	%HbA _{1c}	10.8	8.64 – 13.0
TOSOH HLC723/G7/G8/GX (IFCC)	mmol/mol	94.5	75.6 – 113
Trin Bio CLC385/PDQ/Ultra 2 (DCCT / NGSP)	%HbA _{1c}	11.0	8.80 – 13.2
Trin Bio CLC385/PDQ/Ultra 2 (IFCC)	mmol/mol	96.7	77.4 – 116
Trinity Biotech Tri-stat (DCCT / NGSP)	%HbA _{1c}	12.3	9.84 – 14.8
Trinity Biotech Tri-stat (IFCC)	mmol/mol	111	88.8 – 133
Trinity/Menarini Premier Hb9210 (DCCT / NGSP)	%HbA _{1c}	11.1	8.88 – 13.3
Trinity/Menarini Premier Hb9210 (IFCC)	mmol/mol	97.8	78.2 – 117

Total Haemoglobin		Target	Range
Randox Rx Series	g/dl	12.9	10.3 – 15.5

08 Sep '17 ne