

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

| DATE OF MANUFACTURE: | 2017-01                    |
|----------------------|----------------------------|
| KIT DESCRIPTION:     | BLOOD GAS CONTROL LEVEL II |
| CATALOGUE NO:        | BG5002                     |
| BATCH NO:            | 170BG                      |
| EXPIRY DATE:         | 2018-08                    |

| SAMPLE | ANALYTE  | TARGET | RANGE        | RESULT | UNITS  |
|--------|----------|--------|--------------|--------|--------|
| 170BG  | SODIUM   | 140    | 133 - 147    | 144    | mmol/L |
| 170BG  | CHLORIDE | 96.3   | 88.6 – 104.0 | 92.8   | mmol/L |

APPROVED BY:

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com