



CERTIFICATE OF ANALYSIS

This is to certify that the following kits were tested by
Randox Testing Services and approved by the QC Department.

DATE OF MANUFACTURE: October 2014
KIT DESCRIPTION: Immunoassay Control Level 3
CATALOGUE NO: IA2640
LOT NUMBERS: 1424EC
EXPIRY DATE: 2017-05
SYSTEM OF ANALYSIS: Siemens Immulite 2000

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
Cortisol (nmol/l)	1424EC	731 - 1219	975	899	
				888	
FSH (U/l)	1424EC	41.8 - 62.6	52.2	57.2	
				56.1	

QC TESTED BY: Misha Piracha
DATE: 02-Oct-14
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *Adèle Martin*

Randox Laboratories Ltd. Ardmore, Diamond Road, Crumlin, Co. Antrim, United Kingdom, BT29 4QY
Tel : +44 2894 422413 Fax : +44 2894 452912

FORM NO. 6521-QC
REVISION (2)
22 OCTOBER 2008