

## **CERTIFICATE OF ANALYSIS**

This is to certify that the following kits were tested by Randox Testing Services and approved by the QC Department.

**DATE OF MANUFACTURE:** December 2015

KIT DESCRIPTION: Immunoassay Premium Plus Level 2

CATALOGUE NO: IA3110

LOT NUMBERS: 1575EC

**EXPIRY DATE**: 2018-11

SYSTEM OF ANALYSIS: Siemens Immulite 2000xpi

## **TEST CONTROLS OR VIALS**

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (kU/l)	1575EC	43.6 - 63.4	54.5	48.2	
				48.0	
DHEAS (umol/l)	1575EC	9.68 - 14.52	12.1	11.8	
				12.8	
HCG (U/I)	1575EC	29.4 - 44.1	36.8	38	
				37	

QC APPROVED BY:		adèle Martin.		
RESULT (PASS/FAIL):		Pass		
DATE:	07-Dec-15			
QC TESTED BY:		Aimee Anderson		

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com