

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: January 2017
KIT DESCRIPTION: IMMUNOASSAY PREMIUM PLUS 1
CATALOGUE NO: IA3109
LOT NUMBERS: 1619EC
EXPIRY DATE: 2019-04-28
SYSTEM OF ANALYSIS: Roche Cobas E411

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
TT3 (nmol/l)	1619EC	1.44 – 2.40	1.92	2.2	N/A
				2.18	
				2.24	
TT4 (nmol/l)	1619EC	30.54 – 56.72	43.63	40.78	N/A
				40.29	
				40.58	
LH (mU/ml)	1619EC	2.68 – 4.02	3.35	3.61	N/A
				3.55	
				3.5	

QC TESTED BY: Samantha Purse
DATE: 12 Jan 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY:

Adele Martin

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE MARI R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15736, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016