

RANDOX

CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: March 2017
KIT DESCRIPTION: Immunoassay Premium Plus Level 3
CATALOGUE NO: IA3111
LOT NUMBERS: 1647EC
EXPIRY DATE: 2019-04-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
AFP (KU/l)	1647EC	157 - 235	196	200	N/A
				201	
FT3 (pmol/l)	1647EC	13.3 - 22.1	17.7	17.8	
				18.4	
TSH (mIU/l)	1647EC	17.0 - 25.6	23.1	22.1	
				21.8	

QC TESTED BY: Davina Graham
DATE: 27-Mar-17
RESULT (PASS/FAIL): PASS
QC APPROVED BY: Adele Martin.

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com I www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAF Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland. Company Number NI15736. VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016