



CERTIFICATE OF ANALYSIS

This is to certify that the following kit was tested by a
Randox Testing Laboratory and approved by the QC Department.

DATE OF MANUFACTURE: October 2016
KIT DESCRIPTION: Immunoassay Premium Plus Level 2
CATALOGUE NO: IA3110
LOT NUMBERS: 1662EC
EXPIRY DATE: 2019-10-28
SYSTEM OF ANALYSIS: Siemens Immulite 2000XPI

TEST CONTROLS OR VIALS

ANALYTE	LOT NUMBER	RANGE	TARGET	RESULT	PRECISION (if applicable)
CEA (ug/l)	1662EC	30.9 - 46.3	38.6	44.7	N/A
				42.5	
FT3 (pmol/l)	1662EC	7.95 - 13.3	10.6	11.7	N/A
				12.0	
SHBG (nmol/l)	1662EC	32.1 - 48.1	40.1	38.3	N/A
				36.5	

QC TESTED BY: Davina Graham
DATE: 20-21 Mar 2017
RESULT (PASS/FAIL): PASS
QC APPROVED BY: *Adele Martin*

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** marketing@randox.com **I** www.randox.com

Directors: Dr S P Fitzgerald BSc PhD DSc (Hons) FIAE Man. R J Fitzgerald, W Fitzgerald OBE, R P Kelly BSc FCA.
Randox Laboratories Limited is a company registered in Northern Ireland, Company Number NI15738, VAT Number GB 353 0304 00.

FORM NO. 6521-QC
REVISION (4)
14 April 2016