

RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2016-04-28
KIT DESCRIPTION:	COAGULATION CONTROL LEVEL 3
CATALOGUE NO:	CG5023
BATCH NO:	245CG
EXPIRY DATE:	2019-04-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
245CG	PROTHROMBIN TIME	24.0	19.2-28.8	23.4	SECS
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY: Yette Millar

QUALITY CONTROL REPRESENTATIVE