

## RANDOX QUALITY CONTROL CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

DATE OF MANUFACTURE:	2017-10
KIT DESCRIPTION:	CRP LIQUID CONTROL LEVEL 2
CATALOGUE NO:	CP2480
BATCH NO:	2807CP
EXPIRY DATE:	2019-04-28

SAMPLE	ANALYTE	TARGET	RANGE	RESULT	UNITS
LEVEL II / 2807CP	CRP	21.6	17.3-25.9	20.5	mg/L
N/A	N/A	N/A	N/A	N/A	N/A

APPROVED BY:

Lisa Ren

QUALITY CONTROL REPRESENTATIVE

Randox Laboratories Limited, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom T +44 (0) 28 9442 2413 F +44 (0) 28 9445 2912 E marketing@randox.com | I www.randox.com